

Certificate of Conformity

issued by Curtis-Straus LLC, a wholly owned subsidiary of BV CPS

Company: Danatronics
Product Tested: ECHO
Testing Date: February 13, March 3, 10, 12, 23, 31, April
2, May 12, 22, 23, and 30, 2015
Report No.: EO3841-2

*A sample of the product as configured in the accompanying test report
has been found to comply with the following standards:*

EMC Emissions:

- CFR 47 FCC Part 15 Subpart B Class A emissions requirements (USA)
- EN 55011:2009/A1:2010 Group 1 Class A ISM emissions requirements (EU)
- AS/NZS CISPR 11:2004 Class A ISM emissions requirements (Australia)
- ICES-001 Issue 4 ISM emissions requirements (Canada)
- EN 61000-3-2:2006/A1:2009/A2:2009 Limits for harmonic current emissions (equipment input current up to and including 16A per phase)
- EN 61000-3-3:2013 Limitation of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current up to and including 16 A


EMC Emissions and Immunity:

- EN 61326-1:2013 EMC requirements for Electrical equipment for measurement, control and laboratory use – General Use

Attested to by the hands and seals:

For Curtis-Straus LLC

For the Manufacturer or Importer



Test Engineer 9/15/15
Date

Date



EMC Project Manager 9/15/15
Date